

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Sample: KM0037 Baird Parker Agar
Batch Number: 00002991
Expiry Date: 03.2022
Date Received: 03.2019
Date Tested: 03.2019
Sample Condition: Satisfactory
Sample Number: 354089

Product sample prepared and tested as PP1230 with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Staphylococcus aureus NCTC 12981/ATCC 25923	78	Black colonies with a narrow opaque halo surrounded by a zone of clearing 1.0-3.0mm	Black colonies with a narrow opaque halo surrounded by a zone of clearing 1.0-3.0mm

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2 .

Selectivity	Test Result	Specification
Escherichia coli NCTC 12241/ATCC 25922	Inhibited	Inhibited
Staphylococcus epidermidis NCTC 13360/ATCC 12228	Inhibited	Inhibited

Physical	Result	Specifications
pH	7.4	7.2 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Pale yellow

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0037 Baird Parker Agar
Batch Number: 00002991
Expiry Date: 03.2022
Date Received: 03.2019
Date Tested: 03.2019
Sample Condition: Satisfactory
Sample Number: 354089

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

All of the results on this certificate of analysis relate only of the samples submitted.



Douglas Cameron
Technical Manager, E&O Diagnostics Ltd