

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Sample: KM0161 Slanetz & Bartley Medium
Batch Number: 00003358
Expiry Date: 09.08.2022
Date Received: 2019-08-09
Date Tested: 2019-08-09
Sample Condition: Satisfactory
Sample Number: 361600

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Enterococcus faecalis NCTC 12697/ATCC 29212	114	Maroon colonies	Maroon colonies
Enterococcus faecium NCTC 12202	88	Maroon colonies	Maroon colonies
Enterococcus faecium DSM 2146/ATCC 6057	66	Maroon colonies	Maroon colonies
Enterococcus faecium DSM 20477/ATCC 19434	83	Maroon colonies	Maroon colonies

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2 .

Selectivity	Test Result	Specification
Escherichia coli NCTC 12241/ATCC 25922	Inhibited	Inhibited
Staphylococcus aureus NCTC 12981/ATCC 25923	Inhibited	Inhibited

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Diagnostics Ltd

Sample: KM0161 Slanetz & Bartley Medium
Batch Number: 00003358
Expiry Date: 09.08.2022
Date Received: 2019-08-09
Date Tested: 2019-08-09
Sample Condition: Satisfactory
Sample Number: 361600

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Physical	Result	Specifications
pH	7.2	7.2 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Rose98-8 - 108-9

All of the results on this certificate of analysis relate only of the samples submitted.

Douglas Cameron
Technical Manager, E&O Diagnostics Ltd