

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



E&O Laboratories Ltd

**Sample:** BM2200 - MMGM (SS)  
**Batch Number:** 04088319  
**Expiry Date:** 2021-04-27  
**Date Received:** 2020-10-28  
**Date Tested:** 2020-10-28  
**Sample Condition:** Satisfactory  
**Sample Number:** 389419

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

### Accredited Test Method: See table below

Productivity	Result	Colonial Appearance	Colonial Appearance Specification
Pseudomonas aeruginosa NCTC 12903	Growth	Broth: Acid production = -ve (Purple) Gas production = -ve MacConkey plate: Yellow colonies	Broth: Acid production = -ve (Purple) Gas production = -ve MacConkey plate: Yellow colonies
Enterobacter aerogenes NCTC 10006	Growth	Broth: Acid production = +ve (Yellow) Gas production = -ve MacConkey plate: Red colonies	Broth: Acid production = +ve (Yellow) Gas production = -ve MacConkey plate: Red colonies
Escherichia coli NCTC 12241	Growth	Broth: Acid production = +ve (Yellow) Gas production = +ve MacConkey plate: Red colonies	Broth: Acid production = +ve (Yellow) Gas production = +ve MacConkey plate: Red colonies

For solid media the RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media /  $\geq 50\%$  for selective media. For liquid media growth is determined from a subculture of the test broth inoculated with 10 - 100 cfu of the target organism.

### Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor $\geq 2$ .

Selectivity	Test Result	Specification
Enterococcus faecalis NCTC 12697	Inhibited	Inhibited

Physical	Result	Specifications	Accredited Test Method
Sterility	Conforms	Within acceptable limits	ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C
Sterility sampling is performed in accordance with ISO 2859-1:1999*. The inspection level is $\geq 0.4\%$ of the batch and the reject level $\leq 7$ units depending on batch size.			
pH	6.8	6.7 +/- 0.2	ED/SOP/003 measurement by pH meter
Colour	Conforms	Purple	ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



E&O Laboratories Ltd

**Sample:** BM2200 - MMGM (SS)  
**Batch Number:** 04088319  
**Expiry Date:** 2021-04-27  
**Date Received:** 2020-10-28  
**Date Tested:** 2020-10-28  
**Sample Condition:** Satisfactory  
**Sample Number:** 389419

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Fill Quantity	Conforms	10ml +/- 0.3	ED/SOP/054 by gravimetric determination
---------------	----------	--------------	---

Accredited Test Method for Solid Media	Accredited Test Method for Liquid Media	Accredited Test Method for Diluents
ED/SOP/008 Quantitative evaluation using spread inoculum technique	ED/SOP/008 Semi quantitative inoculation with growth assessment from subculture	ED/SOP/008 Semi quantitative evaluation of viability maintenance after a holding time of 45 minutes

All of the results on this certificate of analysis relate only to the samples submitted.

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications

\*Sterility sampling is outwith scope of accreditation.



Douglas Cameron  
Technical Manager, E&O Laboratories Ltd