

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0168 Triple Sugar Iron Agar
Batch Number: 00003115
Expiry Date: 05.2022
Date Received: 23.05.19
Date Tested: 23.05.19
Sample Condition: Satisfactory
Sample Number: 357357

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Product sample prepared and tested as BM0440 with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
<i>Pseudomonas aeruginosa</i> NCTC 12903/ATCC 27853	conforms	Red slope/Red butt Gas -ve, H2S -ve	Red slope/Red butt Gas -ve, H2S -ve
<i>Escherichia coli</i> NCTC 12241/ATCC 25922	conforms	Yellow slope/Yellow butt Gas , H2S -ve	Yellow slope/Yellow butt Gas , H2S -ve
<i>Salmonella typhimurium</i> NCTC 12023/ATCC 14028	conforms	Red slope/Yellow butt Gas , H2S	Red slope/Yellow butt Gas , H2S
<i>Shigella sonnei</i> NCTC 8574	conforms	Red slope/Yellow butt Gas -ve, H2S -	Red slope/Yellow butt Gas -ve, H2S -
<i>Shigella flexneri</i> ATCC 12022	conforms	Red slope/Yellow butt Gas -ve, H2S -	Red slope/Yellow butt Gas -ve, H2S -
<i>Proteus mirabilis</i> NCTC 10975	conforms	Red slope/Yellow butt Gas , H2S	Red slope/Yellow butt Gas , H2S

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0168 Triple Sugar Iron Agar
Batch Number: 00003115
Expiry Date: 05.2022
Date Received: 23.05.19
Date Tested: 23.05.19
Sample Condition: Satisfactory
Sample Number: 357357

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Physical	Result	Specifications
pH	7.2	7.4 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Red/orange

All of the results on this certificate of analysis relate only of the samples submitted.



Douglas Cameron
Technical Manager, E&O Diagnostics Ltd