

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Sample: KM0084 Bacillus cereus Agar (MYP)
Batch Number: 00004641
Expiry Date: 03.09.2024
Date Received: 08.09.2021
Date Tested: 08.09.2021
Sample Condition: Satisfactory
Sample Number: 408867

Product sample prepared and tested as PP0050 MYP with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Bacillus cereus NCTC 10320	106	Pink colonies with white halo	Pink colonies with white halo
Bacillus subtilis NCTC 10400/ATCC 6633	93	Yellow colonies	Yellow colonies

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2 .		
Selectivity	Test Result	Specification
Escherichia coli NCTC 12241/ATCC 25922	Inhibited	Inhibited

Physical	Result	Specifications
pH	7.0	7.2 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Peach86-4 - 86-7

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0084 Bacillus cereus Agar (MYP)
Batch Number: 00004641
Expiry Date: 03.09.2024
Date Received: 08.09.2021
Date Tested: 08.09.2021
Sample Condition: Satisfactory
Sample Number: 408867

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

All of the results on this certificate of analysis relate only of the samples submitted.



Douglas Cameron
Technical Manager, E&O Diagnostics Ltd