

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Sample: KM0038 Brilliant Green Agar
Batch Number: 00003459
Expiry Date: 30.11.2022
Date Received: 04.11.2019
Date Tested: 04.11.2019
Sample Condition: Satisfactory
Sample Number: 366734

Product sample prepared and tested as PP0060 Brilliant Green Agar with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Salmonella enteritidis NCTC 12694/ATCC 13076	79	Red/pink colonies	Red/pink colonies
Salmonella typhimurium NCTC 12023/ATCC 14028	107	Red/pink colonies	Red/pink colonies

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2 .		
Selectivity	Test Result	Specification
Escherichia coli ATCC 8739	Inhibited	Inhibited
Escherichia coli NCTC 12241/ATCC 25922	Inhibited	Inhibited
Enterococcus faecalis NCTC 12697/ATCC 29212	Inhibited	Inhibited

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0038 Brilliant Green Agar
Batch Number: 00003459
Expiry Date: 30.11.2022
Date Received: 04.11.2019
Date Tested: 04.11.2019
Sample Condition: Satisfactory
Sample Number: 366734

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Physical	Result	Specifications
pH	7.0	6.9 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Tan43-1 - 46-1

All of the results on this certificate of analysis relate only of the samples submitted.



Douglas Cameron
Technical Manager, E&O Diagnostics Ltd