

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0027 MacConkey Agar No. 3
Batch Number: 00003455
Expiry Date: 27.11.2022
Date Received: 29.11.2019
Date Tested: 29.11.2019
Sample Condition: Satisfactory
Sample Number: 368271

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP1720 with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Escherichia coli NCTC 12241/ATCC 25922	65	Red colonies	Red colonies
Pseudomonas aeruginosa NCTC 12903/ATCC 27853	67	Yellow colonies	Yellow colonies
Proteus mirabilis NCTC 10975	71	Yellow colonies	Yellow colonies
Salmonella typhimurium NCTC 12023/ATCC 14028	61	Yellow colonies	Yellow colonies

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2 .

Selectivity	Test Result	Specification
Enterococcus faecalis NCTC 12697/ATCC 29212	Inhibited	Inhibited

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0027 MacConkey Agar No. 3
Batch Number: 00003455
Expiry Date: 27.11.2022
Date Received: 29.11.2019
Date Tested: 29.11.2019
Sample Condition: Satisfactory
Sample Number: 368271

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Physical	Result	Specifications
pH	7.3	7.1 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Pale red with slight violet tinge99-3 - 99-5

All of the results on this certificate of analysis relate only of the samples submitted.



Douglas Cameron
Technical Manager, E&O Diagnostics Ltd