

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** KM0201 Long and Hammer Agar  
**Batch Number:** 00004932  
**Expiry Date:** 25.02.2025  
**Date Received:** 02.03.2022  
**Date Tested:** 02.03.2022  
**Sample Condition:** Satisfactory  
**Sample Number:** 419957

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Product sample prepared and tested as PP3602 Long and Hammer Agar with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Vibrio cholerae NCTC 11348	90	Cream colonies	Cream colonies
Vibrio furnissii NCTC 11218	119	Cream colonies	Cream colonies
Vibrio parahaemolyticus NCTC 10885	120	Cream colonies	Cream colonies
Photobacterium angustum EG0152	72	Cream colonies	Cream colonies

Physical	Result	Specifications
pH	7.2	7.0 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Straw2-8 - 6-8

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



E&O Diagnostics Ltd

**Sample:** KM0201 Long and Hammer Agar  
**Batch Number:** 00004932  
**Expiry Date:** 25.02.2025  
**Date Received:** 02.03.2022  
**Date Tested:** 02.03.2022  
**Sample Condition:** Satisfactory  
**Sample Number:** 419957

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

All of the results on this certificate of analysis relate only of the samples submitted.

**Douglas Cameron**  
Technical Manager, E&O Diagnostics Ltd