

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



E&O Laboratories Ltd

**Sample:** PP0320 - X.L.D. Agar  
**Batch Number:** 04160863  
**Expiry Date:** 2022-08-05  
**Date Received:** 2022-05-30  
**Date Tested:** 2022-05-30  
**Date of Issue:** 2022-06-03  
**Sample Condition:** Satisfactory  
**Sample Number:** 425663

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

The RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media /  $\geq 50\%$  for selective media.

Accredited Test Method: ED/SOP/008 Quantitative evaluation using spread inoculum technique (solid media)

| Productivity                         | RGI (%) | Colonial Appearance                    | Colonial Appearance Specification      |
|--------------------------------------|---------|----------------------------------------|----------------------------------------|
| Shigella sonnei<br>NCTC 8574         | 51      | Red colonies                           | Red colonies                           |
| Salmonella typhimurium<br>NCTC 12023 | 72      | Translucent colonies with black centre | Translucent colonies with black centre |

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor  $\geq 2$ .

| Selectivity                         | Test Result | Specification |
|-------------------------------------|-------------|---------------|
| Escherichia coli<br>NCTC 12241      | Inhibited   | Inhibited     |
| Enterococcus faecalis<br>NCTC 12697 | Inhibited   | Inhibited     |

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**E&O Laboratories Ltd**

**Sample:** PP0320 - X.L.D. Agar  
**Batch Number:** 04160863  
**Expiry Date:** 2022-08-05  
**Date Received:** 2022-05-30  
**Date Tested:** 2022-05-30  
**Date of Issue:** 2022-06-03  
**Sample Condition:** Satisfactory  
**Sample Number:** 425663

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

| Physical                                                                                                                                                                            | Result   | Specifications           | Accredited Test Method                                                                            |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------|--------------------------|---------------------------------------------------------------------------------------------------|
| Sterility                                                                                                                                                                           | Conforms | Within acceptable limits | ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C |
| Sterility sampling is performed in accordance with ISO 2859-1:1999*. The inspection level is $\geq 0.4\%$ of the batch and the reject level $\leq 7$ units depending on batch size. |          |                          |                                                                                                   |
| pH                                                                                                                                                                                  | 7.2      | 7.4 +/- 0.2              | ED/SOP/003 measurement by pH meter                                                                |
| Colour                                                                                                                                                                              | Conforms | 60-1 - 78-1 Red          | ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide            |
| Fill Quantity                                                                                                                                                                       | Conforms | 18-22ml                  | ED/SOP/054 by gravimetric determination                                                           |
| Moisture                                                                                                                                                                            | Conforms | 95% +/- 3%               | ED/SOP/053 with drying and gravimetric determination                                              |

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**E&O Laboratories Ltd**

**Sample:** PP0320 - X.L.D. Agar  
**Batch Number:** 04160863  
**Expiry Date:** 2022-08-05  
**Date Received:** 2022-05-30  
**Date Tested:** 2022-05-30  
**Date of Issue:** 2022-06-03  
**Sample Condition:** Satisfactory  
**Sample Number:** 425663

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

All of the results on this certificate of analysis relate only to the samples submitted.

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications

\*Sterility sampling is outwith scope of accreditation.



1894

**Douglas Cameron**  
Technical Manager, E&O Laboratories Ltd