

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Ref.	Batch	Expiration date
ComASP® Cefiderocol 0.008- 128	75009	021524508	2024.08.03

Kit Content	Batch	Expiration date
ComASP Cefiderocol	020524002	2024.08.03
Iron Depleted - Mueller Hinton Broth	020224067	2024.09.29

ComASP Cefiderocol (empty panel)	030223104
----------------------------------	-----------

Microbiological Performance
Incubation: Ambient air, 16-20 h, 36 ± 2°C

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Expected Results MIC range (mg/L)	Results MIC (mg/L)
<i>Escherichia coli</i> ATCC® 25922	0.06 – 0.5	0.5
<i>Pseudomonas aeruginosa</i> ATCC® 27853	0.06 – 0.5	0.25

Batch Release

Approved

Date 16.02.2024

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.