

Liofilchem®

Certificate of Analysis

Product	Ref.	Batch	Expiration date
ComASP® Cefiderocol 0.008- 128	75009	072224515	2024.12.15

Kit Content	Batch	Expiration date
ComASP Cefiderocol	070824003	2025.01.04
Iron Depleted - Mueller Hinton Broth	041924057	2024.12.15

ComASP Cefiderocol (empty panel)	080123085
----------------------------------	-----------

Microbiological Performance
Incubation: Ambient air, 16-20 h, 36 ± 2°C

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Expected Results MIC range (mg/L)	Results MIC (mg/L)
<i>Escherichia coli</i> ATCC® 25922	0.06 – 0.5	0.25
<i>Pseudomonas aeruginosa</i> ATCC® 27853	0.06 – 0.5	0.25

Batch Release

Approved

Date

23.07.2024

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.