

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Ref.	Batch	Expiration date
AD Fosfomicin 0.25-256 µg/mL	77001 – 77061	072924091	2024.11.26

Microbiological Performance

Incubation: Ambient air, 16-20 h, 35 ± 2°C

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Expected Results MIC range (mg/L)	Results MIC (mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	0.5–4*	2
<i>Enterococcus faecalis</i> ATCC® 29212	32–128	32
<i>Escherichia coli</i> ATCC® 25922	0.5–2*	1
<i>Pseudomonas aeruginosa</i> ATCC® 27853	2–8*	4

*Validated by EUCAST

Batch Release

Approved

Date

21.08.2024

Signature

Quality Control

(This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.