

 S GR11.6 DOC 25 REV09 OF 02/05/2024	<b>CERTIFICATE OF STERILITY AND QUALITY ASSURANCE</b>	N° <b>25-CRT04053</b>
---	---	--------------------------

CODE	LOT N°	EXP. DATE
338C	251239100	28/01/2027

DEVICE DESCRIPTION:	Medium For Virus, Chlamydia, Mycoplasma
---------------------	---

1. Appearance: The medium is clear and light rose-orange in color.
2. Manufacture date: 28/07/2025
3. pH of the medium at 25°C acceptability limits  $7,3 \pm 0,2$ : passed
4. USP Membrane Filtration Sterility Test: Passed
5. Culture Response:

Swabs were inoculated with the test organisms and placed in the UTM medium. Good recovery of all organisms was obtained after 48 hours at 2-6°C and 20-24°C.

Test Organisms

Herpes Simplex Virus Type 2	ATCC VR-734
Influenza A	ATCC VR-1679
Cytomegalovirus	ATCC VR-977
Coxsackievirus B1	ATCC VR-28
<i>Chlamydia trachomatis</i>	ATCC VR-880
<i>Mycoplasma pneumoniae</i>	ATCC 15531

All reported results were obtained following internal SOP instructions.

Goods has been manufactured according to the ISO 13485:2016 system.

By signing below, Copan Italia SpA certifies that the above quality requirements have been met and all information stated above is correct.

01/08/2025

Certificate Date of Issue



Quality Representative  
Copan Italia SpA  
Via F.Perotti 10 – 25125, Brescia - Italy