

PRODUCT CODE and DESCRIPTION:

WR090-01-0011 Colibri™ PRIMARY TUBE WITH SALINE

LOT #	EXPIRATION DATE	QUANTITY
2515200	2026/10/20	69.300,00

Product transparency verification

- Clear and transparent at Visual Inspection: *passed*
- Optical Density Response by nephelometer, in acceptable range: *passed*.

All reported results were obtained at time of Lot release.

COPAN WASP S.r.l. certifies that product quality requirements have been met, information stated above is correct, and that the Manufacturer Quality Management System is compliant to the standards EN ISO 13485:2016 Medical devices - Quality management systems – Requirements for regulatory purposes and EN ISO 9001:2015 Quality Management System – Requirements and following amendments.

2025/11/20

Certificate Date of Issue
(yyyy/mm/dd)



Chief Quality Officer
COPAN WASP S.r.l.
Via A. Grandi, 32
25125 – Brescia
Italy

Tel. +39 030 2687218 | fax +39 030 5541880 | wasp@copangroup.com