

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland, FK4 2HH



E&O Laboratories Ltd

Sample: KM0201 Long and Hammer Agar
Batch Number: 00020539
Expiry Date: 12/11/2028
Date Received: 24/11/2025
Date Tested: 24/11/2025
Date of Issue: 26/11/2025
Sample Condition: Satisfactory

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP3602 Long and Hammer Agar

For Solid Media the RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media / $\geq 50\%$ for a selective media.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
V.cholerae NCTC 11348	116	Cream colonies	Cream colonies
V.furnissii NCTC 11218	114	Cream colonies	Cream colonies
V.paraohaemolyticus NCTC 10885	114	Cream colonies	Cream colonies

Physical	Result	Specification	Test Method
Colour	Conforms	Straw. 2-8 - 6-8	ED/SOP/009 by visual observation. Range measured using Pantone guide.
pH	7.2	7.0 ± 0.2	ED/SOP/003 measurement by pH meter.

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland, FK4 2HH



E&O Laboratories Ltd

Sample: KM0201 Long and Hammer Agar
Batch Number: 00020539
Expiry Date: 12/11/2028
Date Received: 24/11/2025
Date Tested: 24/11/2025
Date of Issue: 26/11/2025
Sample Condition: Satisfactory

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

All of the results on this certificate of analysis relate only to the samples submitted. Test specifications are based on ISO 11133:2014/Amd.2:2020 and internal product specifications



Douglas Cameron
Technical Manager, E&O Laboratories Ltd