

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



E&O Laboratories Ltd

**Sample:** PP3010-P090 COLOREX Candida

**Batch Number:** 00026497

**Expiry Date:** 30/06/2026

**Date Received:** 22/04/2026

**Date Tested:** 22/04/2026

**Date of Issue:** 27/04/2026

**Sample Condition:** Satisfactory

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

The RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media /  $\geq 50\%$  for selective media.

Accredited Test Method: ED/SOP/008 Quantitative evaluation using spread inoculum technique (solid media)

| Productivity           | RGI (%) | Colonial Appearance | Colonial Appearance Specification |
|------------------------|---------|---------------------|-----------------------------------|
| C.albicans NCPF 3179   | 84      | Green colonies      | Green colonies                    |
| C.glabrata NCPF 3309   | 90      | Pink colonies       | Pink colonies                     |
| C.tropicalis NCPF 3290 | 78      | Blue colonies       | Blue colonies                     |

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum  
for selective agar and fluid media. Inhibition indicates selectivity factor  $\geq 2$ .

| Selectivity       | Test Result | Specification |
|-------------------|-------------|---------------|
| E.coli NCTC 12241 | Inhibited   | Inhibited     |

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



E&O Laboratories Ltd

**Sample:** PP3010-P090 COLOREX Candida

**Batch Number:** 00026497

**Expiry Date:** 30/06/2026

**Date Received:** 22/04/2026

**Date Tested:** 22/04/2026

**Date of Issue:** 27/04/2026

**Sample Condition:** Satisfactory

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

| Physical                                                                                                                                                                           | Result   | Specifications           | Accredited Test Method                                                                            |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------|--------------------------|---------------------------------------------------------------------------------------------------|
| Sterility                                                                                                                                                                          | Conforms | Within acceptable limits | ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C |
| Sterility sampling is performed in accordance with ISO 2859-1:1999*. The inspection level is $\geq 0.4\%$ of the batch and the reject level $\leq 7$ units depending on batch size |          |                          |                                                                                                   |
| pH                                                                                                                                                                                 | 6.3      | $6.1 \pm 0.2$            | ED/SOP/003 measurement by pH meter                                                                |
| Colour                                                                                                                                                                             | Conforms | Clear pale straw         | ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide            |
| Fill Quantity                                                                                                                                                                      | Conforms | 18-22ml                  | ED/SOP/054 by gravimetric determination                                                           |
| Moisture                                                                                                                                                                           | Conforms | $96 \pm 1\%$             | ED/SOP/053 with drying and gravimetric determination                                              |

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**E&O Laboratories Ltd**

**Sample:** PP3010-P090 COLOREX Candida

**Batch Number:** 00026497

**Expiry Date:** 30/06/2026

**Date Received:** 22/04/2026

**Date Tested:** 22/04/2026

**Date of Issue:** 27/04/2026

**Sample Condition:** Satisfactory

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

All of the results on this certificate of analysis relate only to the samples submitted.

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications.

\*Sterility sampling is outwith scope of accreditation.



1894

**Douglas Cameron**  
Technical Manager, E&O Laboratories Ltd