Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH

E&O Laboratories Ltd

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Sample: PP3095-P090 COLOREX mSuperCARBA

Batch Number: 00031666

Expiry Date: 06/10/2026

Date Received: 02/09/2026

Date Tested: 02/09/2026

Date of Issue: 07/09/2026

Sample Condition: Satisfactory

The RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media / ≥ 50% for selective media.
Accredited Test Method: ED/SOP/008 Quantitative evaluation using spread inoculum technique (solid media)

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
E.coli NCTC 13476	72	Red colonies	Red colonies
K.pneumoniae NCTC 13442 (OXA-48)	83	Metallic blue colonies	Metallic blue colonies

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2.

Selectivity	Test Result	Specification
E.cloacae NCTC 13406	Inhibited	Inhibited
E.faecalis NCTC 12697	Inhibited	Inhibited
E.coli NCTC 13441 (ESBL)	Inhibited	Inhibited

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH

E&O Laboratories Ltd

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Sample: PP3095-P090 COLOREX mSuperCARBA

Batch Number: 00031666

Expiry Date: 06/10/2026

Date Received: 02/09/2026

Date Tested: 02/09/2026

Date of Issue: 07/09/2026

Sample Condition: Satisfactory

Physical	Result	Specifications	Accredited Test Method
Sterility	Conforms	Within acceptable limits	ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C
Sterility sampling is performed in accordance with ISO 2859-1:1999*. The inspection level is ≥ 0.4% of the batch and the reject level ≤ 7 units depending on batch size			
pH	7.2	7.2 ± 0.2	ED/SOP/003 measurement by pH meter
Colour	Conforms	Straw 7499C	ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide
Fill Quantity	Conforms	18-22ml	ED/SOP/054 by gravimetric determination
Moisture	Conforms	96 ± 1%	ED/SOP/053 with drying and gravimetric determination

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH

E&O Laboratories Ltd

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Sample:	PP3095-P090 COLOREX mSuperCARBA
Batch Number:	00031666
Expiry Date:	06/10/2026
Date Received:	02/09/2026
Date Tested:	02/09/2026
Date of Issue:	07/09/2026
Sample Condition:	Satisfactory

All of the results on this certificate of analysis relate only to the samples submitted.

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications.

*Sterility sampling is outwith scope of accreditation.

UKAS TESTING

1894

Douglas Cameron
Technical Manager, E&O Laboratories Ltd