# Certificate of Analysis

**Client**

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH

E&O Laboratories Ltd

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

**Sample:** PP3056-P090 COLOREX MRSA

**Batch Number:** 00030379

**Expiry Date:** 10/09/2026

**Date Received:** 31/07/2026

**Date Tested:** 31/07/2026

**Date of Issue:** 03/08/2026

**Sample Condition:** Satisfactory

The RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media / ≥ 50% for selective media.
Accredited Test Method: ED/SOP/008 Quantitative evaluation using spread inoculum technique (solid media)

| Productivity | RGI (%) | Colonial Appearance | Colonial Appearance Specification |
|---|---|---|---|
| S.aureus NCTC 10442 | 83 | Pink/mauve colonies | Pink/mauve colonies |

**Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2.**

| Selectivity | Test Result | Specification |
|---|---|---|
| S.aureus NCTC 12981 | Inhibited | Inhibited |
| P.aeruginosa NCTC 12903 | Inhibited | Inhibited |
| S.epidermidis NCTC 13360 | Inhibited | Inhibited |

# Certificate of Analysis

**Client**

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH

E&O Laboratories Ltd

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

**Sample:** PP3056-P090 COLOREX MRSA

**Batch Number:** 00030379

**Expiry Date:** 10/09/2026

**Date Received:** 31/07/2026

**Date Tested:** 31/07/2026

**Date of Issue:** 03/08/2026

**Sample Condition:** Satisfactory

| Physical | Result | Specifications | Accredited Test Method |
|---|---|---|---|
| Sterility | Conforms | Within acceptable limits | ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C |
| Sterility sampling is performed in accordance with ISO 2859-1:1999*. The inspection level is ≥ 0.4% of the batch and the reject level ≤ 7 units depending on batch size | | | |
| pH | 6.8 | 6.9 ± 0.2 | ED/SOP/003 measurement by pH meter |
| Colour | Conforms | Straw 4001C | ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide |
| Fill Quantity | Conforms | 18-22ml | ED/SOP/054 by gravimetric determination |
| Moisture | Conforms | 92 ± 2% | ED/SOP/053 with drying and gravimetric determination |

# Certificate of Analysis

**Client**

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH

**E&O Laboratories Ltd**

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

| | |
|---|---|
| **Sample:** | PP3056-P090 COLOREX MRSA |
| **Batch Number:** | 00030379 |
| **Expiry Date:** | 10/09/2026 |
| **Date Received:** | 31/07/2026 |
| **Date Tested:** | 31/07/2026 |
| **Date of Issue:** | 03/08/2026 |
| **Sample Condition:** | Satisfactory |

All of the results on this certificate of analysis relate only to the samples submitted.

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications.

*Sterility sampling is outwith scope of accreditation.

UKAS TESTING

1894

**Douglas Cameron**
**Technical Manager, E&O Laboratories Ltd**