COPAN S GR11.6 DOC 25 REV09 OF 02/05/2024	**CERTIFICATE OF STERILITY AND QUALITY ASSURANCE**	**N°** **26-CRT04301**

CODE	LOT N°	EXP. DATE
355C	261287000	07/04/2028

DEVICE DESCRIPTION:	Medium + Regular Flocked Swab

1. Appearance: The medium is clear and light rose-orange in color.

2. Manufacture date: 07/10/2026

3. pH of the medium at 25°C acceptability limits 7,3 ± 0,2: passed

4. USP Membrane Filtration Sterility Test: Passed

5. Culture Response:

   Swabs were inoculated with the test organisms and placed in the UTM medium. Good recovery of all organisms was obtained after 48 hours at 2-6°C and 20-24°C.

   Test Organisms

Herpes Simplex Virus Type 2	ATCC VR-734
Influenza A	ATCC VR-1679
Cytomegalovirus	ATCC VR-977
Coxsackievirus B1	ATCC VR-28
Chlamydia trachomatis	ATCC VR-880
Mycoplasma pneumoniae	ATCC 15531

All reported results were obtained following internal SOP instructions.
Goods has been manufactured according to the ISO 13485:2016 system.
By signing below, Copan Italia SpA certifies that the above quality requirements have been met and all information stated above is correct.

15/10/2026
Certificate Date of Issue



Quality Representative
Copan Italia SpA
Via F.Perotti 10 – 25125, Brescia - Italy